

NRPP Performance Test Analysis Reporting Form Continuous Radon Monitor

Company Name: Arundel Home Inspection LLC

Chamber Information

Air Velocity: 15 ± 5 ft/min

Elevation at chamber facility = 820 feet (250 m)

Ambient Gamma Background Radiation = 7 μ R/h**Exposure Data: Times are in 24-hr (military) units, Eastern Standard Time (EST)**

Device Serial #	Beginning of Chamber Exposure	End of Chamber Exposure	Temp.	Relative Humidity
<u>35531016</u>	Date <u>6/15/15</u> Time <u>1153</u>	Date <u>6/17/15</u> Time <u>1153</u>	<u>70.9</u> °F	<u>49.6</u> %

Measurement Result


Monitor Serial # 35531016 Measurement (pCi/L or Bq/m³) 25.9 pCi/L

Date result must be received
by Bowser-Morner, Inc.:

<u>7/1/15</u>

NOTE: Report ONLY ONE measurement value above. If this were a home test, what single value would you report to the homeowner? Report no more than one digit after the decimal. Do not send a printout from the monitor.

Signature



Date

6-28-2015

Mail: P.O. Box 51, Dayton, OH 45401, Fax: (937) 233-2024 or email: radon@bowser-morner.com this form back to Bowser-Morner with result. **Please allow a minimum of one week for response.**